



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 23-001315/D005.R002
Report Date: 01/18/2024
ORELAP#: OR100028
Purchase Order:
Received: 01/31/23 13:42

This is an amended version of report# 23-001315/D005.R001.
 Reason: Updated customer information.

Customer: Door County Cannabis Co.
Product identity: Live Delta 8 Extract- Wesley's Wish
Client/Metric ID: 2301072
Laboratory ID: 23-001315-0002

Summary

Potency:

Analyte	Result (%)		
Δ8-THC	72.6		CBD-Total 3.07%
CBD-A	3.35		THC-Total 0.215%
THC-A	0.245		(Reported in percent of total sample)
CBD	0.132		
CBG-A	0.103		



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 23-001315/D005.R002
Report Date: 01/18/2024
ORELAP#: OR100028
Purchase Order:
Received: 01/31/23 13:42

Customer: Door County Cannabis Co.

Product identity: Live Delta 8 Extract- Wesley's Wish
Client/Metric ID: 2301072
Sample Date:
Laboratory ID: 23-001315-0002
Evidence of Cooling: No
Temp: 15 °C
Relinquished by: Ramos

Sample Results

Potency	Method: J AOAC 2015 V98-6 (mod) ^p			Units %	Batch: 2301072	Analyze: 2/2/23 11:20:00 PM
Analyte	As Received	Dry weight	LOQ	Notes		
CBD	0.132		0.0739			
CBD-A	3.35		0.0739			
CBD-Total	3.07		0.139			
CBG	< LOQ		0.0739			
CBG-A	0.103		0.0739			
CBG-Total	< LOQ		0.138			
CBN	< LOQ		0.0739			
Δ10-THC-9R	< LOQ		0.0739			
Δ10-THC-9S	< LOQ		0.0739			
Δ10-THC-Total	< LOQ		0.148			
Δ8-THC	72.6		0.739			
Δ9-THC	< LOQ		0.0739			
THC-A	0.245		0.0739			
THC-Total	0.215		0.139			
Total Cannabinoids	76.4					



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 23-001315/D005.R002
Report Date: 01/18/2024
ORELAP#: OR100028
Purchase Order:
Received: 01/31/23 13:42

These test results are representative of the individual sample selected and submitted by the client.

Abbreviations

Limits: Action Levels per OAR-333-007-0400, OAR-333-007-0210, OAR-333-007-0220, CCR title 16-division 42. BCC-section 5723

Limit(s) of Quantitation (LOQ): The minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence.

^b = ISO/IEC 17025:2017 accredited method.

Units of Measure

% = Percentage of sample

% wt = µg/g divided by 10,000

Approved Signatory

Derrick Tanner
General Manager



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 23-001315/D005.R002
Report Date: 01/18/2024
ORELAP#: OR100028
Purchase Order:
Received: 01/31/23 13:42

Revision: 1 Document ID: 7148
 Legacy ID: Worksheet Validated 04/20/2021

Laboratory Quality Control Results

JAOAC2015 V986 Batch ID: 2301072

Laboratory Control Sample									
Analyte	LCS	Result	Spike	Units	% Rec	Limits		Evaluation	Notes
CBDA	2	0.103	0.100	%	103	80.0	- 120	Acceptable	
CBDV	2	0.113	0.100	%	113	80.0	- 120	Acceptable	
CBE	2	0.106	0.100	%	106	80.0	- 120	Acceptable	
CBDA	1	0.112	0.111	%	101	90.0	- 110	Acceptable	
CBSA	1	0.112	0.110	%	101	80.0	- 120	Acceptable	
CBS	1	0.117	0.115	%	102	80.0	- 120	Acceptable	
CB	1	0.114	0.113	%	100	90.0	- 110	Acceptable	
THCV	2	0.100	0.100	%	100	80.0	- 120	Acceptable	
δ8THCV	2	0.104	0.100	%	104	80.0	- 120	Acceptable	
THCVA	2	0.100	0.100	%	100	80.0	- 120	Acceptable	
CBN	1	0.117	0.118	%	99.4	80.0	- 120	Acceptable	
exo-THC	2	0.0976	0.100	%	97.6	80.0	- 120	Acceptable	
δ9THC	1	0.116	0.119	%	98.1	90.0	- 110	Acceptable	
δ8THC	1	0.118	0.120	%	98.7	90.0	- 110	Acceptable	
9Sδ10THC	1	0.0589	0.060	%	97.4	80.0	- 120	Acceptable	
CB	2	0.0964	0.100	%	96.4	80.0	- 120	Acceptable	
9Rδ10THC	1	0.0533	0.055	%	97.3	80.0	- 120	Acceptable	
CB	2	0.103	0.100	%	103	80.0	- 120	Acceptable	
THCA	1	0.105	0.108	%	97.1	90.0	- 110	Acceptable	
CBSA	2	0.104	0.100	%	104	80.0	- 120	Acceptable	
CBLA	2	0.104	0.100	%	104	80.0	- 120	Acceptable	
CB	2	0.100	0.100	%	100	80.0	- 120	Acceptable	

Method Blank						
Analyte	Result	LOQ	Units	Limits	Evaluation	Notes
CBDA	<LOQ	0.077	%	< 0.077	Acceptable	
CBDV	<LOQ	0.077	%	< 0.077	Acceptable	
CBE	<LOQ	0.077	%	< 0.077	Acceptable	
CBDA	<LOQ	0.077	%	< 0.077	Acceptable	
CBSA	<LOQ	0.077	%	< 0.077	Acceptable	
CBS	<LOQ	0.077	%	< 0.077	Acceptable	
CB	<LOQ	0.077	%	< 0.077	Acceptable	
THCV	<LOQ	0.077	%	< 0.077	Acceptable	
δ8THCV	<LOQ	0.077	%	< 0.077	Acceptable	
THCVA	<LOQ	0.077	%	< 0.077	Acceptable	
CBN	<LOQ	0.077	%	< 0.077	Acceptable	
exo-THC	<LOQ	0.077	%	< 0.077	Acceptable	
δ9THC	<LOQ	0.077	%	< 0.077	Acceptable	
δ8THC	<LOQ	0.077	%	< 0.077	Acceptable	
9Sδ10THC	<LOQ	0.077	%	< 0.077	Acceptable	
CB	<LOQ	0.077	%	< 0.077	Acceptable	
9Rδ10THC	<LOQ	0.077	%	< 0.077	Acceptable	
CB	<LOQ	0.077	%	< 0.077	Acceptable	
THCA	<LOQ	0.077	%	< 0.077	Acceptable	
CBSA	<LOQ	0.077	%	< 0.077	Acceptable	
CBLA	<LOQ	0.077	%	< 0.077	Acceptable	
CB	<LOQ	0.077	%	< 0.077	Acceptable	

Abbreviations

ND - None Detected at or above MRL
 RPD - Relative Percent Difference
 LOQ - Limit of Quantitation

Units of Measure:

% - Percent



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 23-001315/D005.R002
Report Date: 01/18/2024
ORELAP#: OR100028
Purchase Order:
Received: 01/31/23 13:42

Revision: 1 Document ID: 7148
 Legacy ID: Worksheet Validated 04/20/2021

Laboratory Quality Control Results

JAOAC2015 V986		Batch ID: 2301072						
Sample Duplicate		Sample ID: 23-0011020001						
Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Evaluation	Notes
CBDVA	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CBDV	0.430	0.438	0.077	%	1.83	< 20	Acceptable	
CBE	1.31	1.34	0.077	%	2.06	< 20	Acceptable	
CBD	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CBSA	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CBS	0.289	0.284	0.077	%	1.70	< 20	Acceptable	
CBD	84.3	85.7	0.077	%	1.59	< 20	Acceptable	
THCV	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
δ8THCV	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
THCVA	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CBN	0.314	0.321	0.077	%	2.34	< 20	Acceptable	
exo-THC	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
δ9THC	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
δ8THC	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
9Sa10THC	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CB	0.0921	0.0937	0.077	%	1.72	< 20	Acceptable	
9Ra10THC	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CB	0.398	0.407	0.077	%	2.20	< 20	Acceptable	
THCA	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CBSA	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CBLA	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CB	1.09	1.10	0.077	%	0.898	< 20	Acceptable	

Abbreviations

ND - None Detected at or above MRL
 RPD - Relative Percent Difference
 LOQ - Limit of Quantitation

Units of Measure:

% - Percent



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 23-001315/D005.R002
Report Date: 01/18/2024
ORELAP#: OR100028
Purchase Order:
Received: 01/31/23 13:42





Explanation of QC Flag Comments:

Code	Explanation
Q	Matrix interferences affecting spike or surrogate recoveries.
Q1	Quality control result biased high. Only non-detect samples reported.
Q2	Quality control outside QC limits. Data considered estimate.
Q3	Sample concentration greater than four times the amount spiked.
Q4	Non-homogenous sample matrix, affecting RPD result and/or % recoveries.
Q5	Spike results above calibration curve.
Q6	Quality control outside QC limits. Data acceptable based on remaining QC.
R	Relative percent difference (RPD) outside control limit.
R1	RPD non-calculable, as sample or duplicate results are less than five times the LOQ.
R2	Sample replicates RPD non-calculable, as only one replicate is within the analytical range.
LOQ1	Quantitation level raised due to low sample volume and/or dilution.
LOQ2	Quantitation level raised due to matrix interference.
B	Analyte detected in method blank, but not in associated samples.
B1	The sample concentration is greater than 5 times the blank concentration.
B2	The sample concentration is less than 5 times the blank concentration.